

PRODUCT CERTIFICATE

Product name:	Kjeltabs Cu/3.5
Batch/lot no:	AA17 176226
Use before:	OCT 2030
Date of release of this certificate:	09 December, 2025
P/N:	60046515
Application:	Reagent for Kjeltac instruments

PRODUCT TESTING RESULTS:

Class of method	Result	Specification
Appearance	Complies	Green or Blue speckled flat tablets
Average Weight Range	3.91 g	3.75 – 4.05 g
Weight Uniformity	3.84 – 3.96 g	NMT 10% outside 3.70 – 4.10 g None outside 3.60 – 4.20 g
Friability	3.06 %	NMT 5 %
Nitrogen	Tested every 10 th batch	NMT 0.05 %
Copper Sulphate, 5H ₂ O	Tested every 10 th batch	360 – 440 mg/tablet
Diameter	16.6 mm	15.8 – 16.8 mm

Remarks: Storage: At room temperature. Keep container well closed and protected from direct sunlight and moisture.
Normal Transport Conditions: Temperature not critical.

This batch was released by:

Mercy Kingsley
QC-Department



ANALYTICS BEYOND MEASURE

FOSS Analytical A/S
Nils Foss Allé 1
DK-3400 Hillerød
Denmark
Comp.Reg.No. 73399815

Phone: +45 7010 3370
www.fossanalytics.com

Legal notice:

The general legal notice/disclaimer that applies to this certificate is found on our website www.fossanalytics.com.